

POLICE

HEADQUARTERS

J&K

SRINAGAR.

**Subject :- Corrigendum in e-NIT No. 28 of 2016 dated 12.09.2016 floated by this Hqrs. for procurement of Hospital equipments.**

In view of some queries raised by intending firms, e-NIT No. 28 of 2016 dated 12.09.2016 read with Extension Notice issued under endstt. No. Prov-II/HE-01/14-15/58671-700 dated 06.10.2016 and 60631-50 dated 19.10.2016, specifications of following equipment are hereby amended as under:-

S.No.	Name of the item with specifications	Tentative quantity
1.	<p><b><u>PHACO EMULSIFICATION MACHINE</u></b></p> <ul style="list-style-type: none"><li>⇒ 4 to 6 Piezo-electric crystals U/S Titanium Handpiece with Frequency of approx. 37 KHz or above.</li><li>⇒ VISCO mode to remove viscoelastic from the eye.</li><li>⇒ U/S Delivery modes: Continuous, Pulse, Burst and Programmable emission mode (adjustable duty cycle).</li><li>⇒ Continuous tip stroke length control and adjustment for consistent U/S power delivery.</li><li>⇒ Range of Phaco tips like straight, Bent and flared tips, ABS</li><li>⇒ MICS Phaco tips for 2.2 mm incision and above.</li><li>⇒ Inbuilt bipolar Diathermy.</li><li>⇒ Peristaltic aspiration system with vacuum range upto 650 mmHg and flow rate upto 60 cc/min with adjustable aspiration rise time.</li><li>⇒ Autoclavable I/A tubing with facility of Anterior chamber surge protection system.</li><li>⇒ Vitrectomy cutting rate ranging upto 1500 cuts per minute with Pneumatically powered Guillotine Cutter.</li><li>⇒ Color LCD with touch screen and user friendly control software.</li><li>⇒ Programmable single/dual linear foot switch for linear/ panel control of all equipment functions.</li><li>⇒ USFDA/ European CE Approved.</li><li>⇒ Voice confirmation during mode changes.</li><li>⇒ Phaco Trolley with Manually adjustable I/V Stand.</li><li>⇒ Suitable UPS.</li><li>⇒ Accessories:<ul style="list-style-type: none"><li>○ U/S Hand pieces.</li><li>○ Co-axial I/A Hand pieces.</li><li>○ Bimanual I/A Hand pieces.</li><li>○ Phaco Tips.</li><li>○ Silicon Sleeves.</li><li>○ Test Chambers.</li><li>○ I/A Tubings.</li><li>○ Diathermy Forceps.</li><li>○ Diathermy Cable.</li><li>○ Tip Wrench.</li><li>○ Sterilization Tray.</li><li>○ Vitrectomy cutters.</li></ul></li></ul>	01 No.
2.	<p><b><u>OPERATING MICROSCOPE (OPHTHALMOLOGY)</u></b></p> <ul style="list-style-type: none"><li>⇒ Optics:<ul style="list-style-type: none"><li>○ Apochromatic corrected optics with objective F = 200 mm.</li><li>○ Motorized zoom type magnification changer.</li><li>○ Tiltable binocular tubes (0-90°)</li><li>○ Focus range: 30- 45 mm.</li><li>○ Assistant scope with Beam Splitter.</li></ul></li><li>⇒ Illumination:</li></ul>	01 No.

	<ul style="list-style-type: none"> <li>○ LED Light Source.</li> <li>○ Filter: Heat absorbing, Blue correction filter, yellow filter, Retinal Protection.</li> </ul> <p>⇒ Operation:</p> <ul style="list-style-type: none"> <li>○ Floor Stand with lockable wheels – cataract version.</li> <li>○ Motorised fine focus. X-Y movement.</li> <li>○ Continuous Zoom Magnification from foot switch.</li> </ul> <p>⇒ Optional Accessories:</p> <ul style="list-style-type: none"> <li>○ Beam Splitter with Camera and C-mount.</li> <li>○ Video recording facility.</li> </ul>	
3.	<p><b>CAUTERY</b></p> <p>I. The electro surgical unit should be microprocessor controlled. US FDA and European CE &amp; ISO certificate.</p> <p>II. The offered equipment accessories and consumable should be of same offered original equipment manufacturer.</p> <p>III. The electro surgical unit should be micro controller based and adjust the micro to get the desired facility to use under water cutting.</p> <p>IV. It should have facility for upto 09 programmable and facility to recall the last setting used by user and should be able to record last 100 events/ errors.</p> <p>V. System should have following features, modes, inbuilt- Auto cut, electro surgery (ES), cutting with minimal to medium homeostasis/ Dry cut-ES cutting with moderate to intense homeostasis/ forced coag-Effective, fast pinpoint coagulation/ bipolar soft coag., Non sparking.</p> <p>VI. Should have monopolar and bipolar facility with socket for each and two neutral sockets.</p> <p>VII. Should have forced coagulation/ spray coagulation.</p> <p>VIII. Should be covered by comprehensive maintenance contract.</p> <p>IX. The company should provide hassle free maintenance at all times.</p>	01 No.

However, other terms and conditions of the NIT shall remain un-changed.




AIG/Provision/ Transport)  
For Director General of Police,  
J&K, Srinagar.

21/10 Dated:- 22: 10.2016

No:Prov-II/HE-1/14-15/69370-400  
Copy to the:-

1. Principal Secretary to Govt. Home Department J&K Srinagar for favour of information.
2. All IsGP for information.
3. Director General information J&K Srinagar for favour of information and with the request of arranging publication of corrigendum of e.tender in National Dailies viz, The Hindu, Decan Herald, Hindustan Times, Indian Express and in leading local news paper i.e Greater Kashmir in Srinagar and Excelsior in Jammu and Tender Journals etc. for wide publicity. The payment for publication in National Dailies shall be made by the PHQ against verified bills.
4. DIG CKR Srinagar/ DIG J-K Range Jammu
5. Director Industries Jammu/Srinagar for information and n/action.

6. Director Door Darshan Jammu/ Srinagar
7. Station Director Radio Kashmir Srinagar/ Jammu  
,, for putting the corrigendum on AIR for three consecutive days.
8. Manager Govt. Press Srinagar/Jammu for publication of corrigendum of e-tender in Govt. gazette.
9. SSP J&K CID Cell, 11 Harish Chander Mathur Lane Kasturba Gandhi Marg New Delhi-110001 for information and n/action.
10. PRO to DGP J&K Srinagar alongwith 02 copies of the corrigendum of e-tender for publication purpose. He will personally liaise with Directorate Information for publication of the corrigendum immediately and make available concerned Newspaper in which corrigendum is published to AIG (Prov/Tpt) PHQ J&K for record and reference.
11. In-Charge PCS Gulshan Ground Jammu/ Zewan Sgr.
12. S.O e-procurement/In-Charge IT Centre PHQ for placing corrigendum on the website of J&K Police and J&K Govt. Portal.

  
AIG (Provision/ Transport)  
For Director General of Police,  
J&K, Srinagar

21/10