

Website: www.jkpolice.gov.in email: phqjk@jkpolice.gov.in HEADQUARTERS

POLICE KASHMIR JAMMU

Clarifications/Amendment

In response to the queries raised by different firms under etender No 39 of 2020 dated 31.08.2020 floated for procurement of Medical Equipments for Trauma Centre PCR Srinagar, the necessary amendments as per Annexure "A" of this notice are issued for the information of intending bidders.

The intending bidders may visit JK Govt.e-tenders portal on www.jktenders.gov.in for comprehensive amendments/clarifications.

However, other terms and conditions of the eNIT shall remain the same.

> sd/-Rayees Mohammad Bhat (IPS AIG(Provision & Transport), For Director General of Police, J&K- Jammu.

No.Prov-II/Trauma-08-part-ii/2020-21/285/_ 70 Dated: - /5 /01/2021. Copy to the:-

- 1. Principal Secretary to Govt. Home Department J&K Jammu
- 2. All ADGsP J&K Jammu for favour of information
 - 3. All IsGP J&K Police for favour of information.
 - 4. Director Door Darshan Jammu/Srinagar and Station Director All India Radio Srinagar/Jammu for putting the notice on air for three consecutive days.
 - 5. Manager Government Press Srinagar/Jammu for publication of the Notice in Govt Gazette.
 - 6. AIG Communication PHQ J&K Jammu for information with the request to arrange publication of Notice in 02 widely circulated national and 02 State Dailies through information department.
 - 7. FA/CAO PHQ for info. This is in reference to UO No. Acctts/ PHO/2433 dated 26.11.2020.
 - 8. Medical Superintendent, Police Hospital Srinagar for information.
 - 9. §.O e-Procurement PHQ, J&K for information and n/action. .S.O (IT) PHQ, for placing notice on J&K Police website.

AIG (Provision & Transport), For Director General of Police, J&K- Jammu.

Annexure"A" <u>Technical Parameter of MRI Machine.</u>

Sr.	Actual Specification	Amendment requested	Amended contents
1	2		
4	RF System The system should be quoted with Dual RF Amplifier with latest B1 homogeneity compensation software like Multi drive/True shape/Multi Transmit software with small FOV imaging software like Zoom IT/Diffusion Zoom Imaging or equivalent.	The system should be capable of 30 KW or more peak power either through single or Dual RF Amplifier with latest B1 homogeneity compensation software like Multi Drive/True Form/Multi Transmit software with small FOV imaging software like Zoom IT/diffusion Zoom imaging or equivalent	col 3
	It should support parallel acquisition techniques with a factor of upto 4 in 2D	Should allow parallel acquisition technique	Amended as requested in col 3
8.3		/ 32 channel for the Z- Axis max FOV(50 cm) ir	requested in col 3

Technical Parameters of CT Scan Machine

Technicus 1 d. ames					
Sr.	Actual Specification 2	Amendment desired 3	Amended contents		
1	be essentially dual focus with capacity of more than 5 MHU	essentially dual focus	requested in		
16.3	System should incorporate all the features as per the December 2010 RSNA	features as per December 2016 RSNA Standards/declaration	requested in col 3		
BOQ	including consum	quote the rate of comp nables in relevant rows of f MRI & CT Scan Machine ma	BOQ.		



the intending firms in the Miscillinous row I & II respectively of the BOQ.

iii) The rates needs to be quoted by the firms in INR only.

iv) In order to examine the issues pertaining to site modifications, Civil and Electric work of OT and other equipments, PHQ has appointed a committee to examine such requirements to address issues separately.

v) Firms need to offer complete solutions as asked in the NIT any deficit items supplied by the firm will invite penal action and cancellation of contract with the

approved supplier.

vi) Approved supplier shall be fully responsible for the proper functionality of supplied items during warranty period and under AMC on mutually agreed terms and conditions after expiry of warranty peiod.

Rayees Monammad Bhat (IPS)

AIG (Provision & Transport),

For Director General of Police,

J&K Jammu.