



J&K POLICE HEADQUARTERS
JAMMU/SRINAGAR

CORRIGENDUM

Reference: i) Bid nr GEM/2022/B/2563030 dated 21.09.2022, floated for Procurement of Medical Equipment for Trauma Centre PCR Srinagar.

ii) Corrigendum nr Prov-II/Trauma-08-ii/Part-B/2020-21/8445-47 dated 25.09.2023

1. Vide reference-ii, corrigendum to the queries of prospective bidders was uploaded on GeM. Consequent upon uploading of corrigendum on GeM portal, following firms have again furnished their queries/requests to PHQ for clarification/amendment.
2. The queries of the intending bidders were examined by the committee and amendments/clarifications in schedule 01, 02, 04, 07, 10 and 17 of GeM bid are hereby issued as per tables 1 to 6 below for the information of bidders.

Table-1 Corrigendum for Schedule-1 (07 OT Lights LED)

1	2	3	4	5	6	7	8
S nr	Name of firm which raised the query	Bid Document Name	Clause nr/ Name	Original Provision as per bid	1 st reply/ Amendment / Clarification issued vide reference-ii above	Clarification /query/ Amendment required by bidder	Reply/ Amendment / Clarification hereby issued as 2 nd amendment
1	M/S SR TechnoMed Bakshi Nagar Jammu	Bid specifications	Mobile operating light P-1	Should offer color temperature of 4300 K	Should offer color temperature of 43000K-5000K	Fixed Color temperature between 4300K-5000K or adjustable variable colour temperature	No change
			Operation Theatre light P-6	Main dome should have at least 80 LED and satellite dome 60	Main dome should have 90 or more LED, satellite dome should 60 or more	The number of LED should not be specified as with the new LEDs and reflector technology available. One can achieve the required lux with less number of LEDs also. If the number range of LED range can be a	No change

			range of LED range can be ≥ 50 in Major and satellite Dome respectively	
P-10	Depth of illumination should at least 1200 mm or more	Depth of illumination should be 1200mm or more	This range is very specific. Request to amend to 900-1200mm so more companies can offer and we can level playing field	No change
Ceiling light: single Dome P-11	Focal distance($d1+d2$) =0.8 to 1.2 m	Focal distance should be at least range from 0.75 to 1.3m.	Focal distance ($d1+d2$) Normally this description is used for Camera system or do you mean it is $L1+L2$ range of 750mm-1300mm. Is any range within the above is acceptable.	Focal distance should be at least range from 0.75 to 1.3m. However maximum range can be acceptable
P-13	The 10x4 zoom is very less for critical surgeries a higher zooming is required for better visualization of surgical site. With full HD at least 10x optical and 12x Digital is the minimum one should have. Now days there are 30x optical and 12x digital in the market. So, we suggest the camera zoom should be atleast 10x optical and 12 x digital	10x-12x Camera should have $\geq 10x$ optical and digital zoom $\geq 4x$	Camera should have 10 X optical and 12 X digital zoom	No change
P-14	Both domes should have touch screen	-	Since touch screen is very sensitive and Surgeons/Nurses/Technicians	Need to be touch screen